

1.01 / Item

SARS-CoV-2 Antigen Rapid Test (Swab) professional use COD. EG26APT

1.02 / Imported by
Kimesta d.o.o.

1.03 / Origin
China

1.04 / CE Certification
Disponibile a richiesta

1.05 / Intended use
For diagnosis of pathogens

1.06 / EU Regulation
Directive IVDD 98/79/CE

1.07 / Standards
EN ISO 13485:2016, EN ISO 15223-1:2016, EN ISO 14791:2012, EN ISO 13975:2003, EN ISO 18113-1:2011, EN ISO 18113-2:2011, EN ISO 13612:2002/AC:2002, EN ISO 17511:2020, EN ISO 23640:2015, EN ISO 13641:2002, EN ISO 15223-1:2016

1.08 / REF n.
ICOV-502

1.09 / RDM n.
1989153

1.10 / Description
In vitro kit for the qualitative determination of the new antigen coronavirus, using nasopharyngeal and / or throat swabs. Result in 15 minutes. High precision, specificity and sensitivity. Packaging: 20 tests.

1.11 / Instruction for use and storage
Store as packaged in the sealed pouch at room temperature or refrigerated (2-30°C). The test is stable through the expiration date printed on the sealed pouch. The test must remain in the sealed pouch until use. DO NOT FREEZE. Do not use beyond the expiration date.



2.01/ Primary packaging

Quantity	20 pcs
Dimension	200x130x90mm
Weight-material	320gr -Cardboard

2.02 / Secondary packaging

Quantità	480 pezzi
Misure	55x40x46cm
Peso - materiale	13kg- Carrugated cardboard
Pallet	24